

CHOLESTEROL REFERENCE METHOD LABORATORY NETWORK

Certificate of Traceability

This certifies that

***Polymer Technology Systems Diagnostics
Indianapolis,***

has documented traceability to the **National Reference System for Cholesterol** by performing a direct comparison with the cholesterol reference method using fresh human specimens which cover the National Cholesterol Education Program medical decision points. This analytical system is representative of the manufacturer's product and has demonstrated the ability to meet the NCEP's performance criteria for accuracy and

**Among-run CV
0.4%**

**Average Bias
3.0%**

**Total Error
3.8%**

The comparison was performed with

***Northwest Lipid Metabolism and Diabetes Research Laboratories
Seattle, Washington***

The system evaluated was:

Instrument:

***PTS Diagnostics
Cardio Chek® PA®***

Calibrator:

Lot #:

Set Point(s):

Cholesterol reagent:

Matrix:

Whole Blood

Lot #: P632

Date of evaluation: 6/15/2017

Date of expiration: 6/15/2019

CRMLN Laboratory Director

P. Marcovilla 7-18-17