

<b>ChekMate Strips Range Card</b>				<b>Batch #:</b>	<b>0860-31017-01</b>
<i>Please enter ranges listed below on the Results Sheet included with your ChekMate kit.</i>				<b>Expiration date:</b>	<b>2019 05 06</b>
	<b>C1 RED</b>	<b>C2 RED</b>	<b>C3 RED</b>	<b>C1 GREEN</b>	
<b>Level 1</b>	21.50 – 28.00	21.50 – 28.00	20.00 – 27.00	15.00 – 23.00	
<b>Level 2</b>	37.00 – 45.00	37.50 – 45.00	37.75 – 43.75	42.75 – 51.25	
<i>ChekMate results are reported as % reflectance and should not be compared to values obtained with PTS Panels® test strips.</i> PTS Diagnostics Customer Service: 877-870-5610   Fax: 317-870-5608   Outside of US: +1-317-870-5610 Email: customerservice@ptsdiagnostics.com					
				PS-000882 Rev. 2	

<b>ChekMate Strips Range Card</b>				<b>Batch #:</b>	<b>0860-31017-01</b>
<i>Please enter ranges listed below on the Results Sheet included with your ChekMate kit.</i>				<b>Expiration date:</b>	<b>2019 05 06</b>
	<b>C1 RED</b>	<b>C2 RED</b>	<b>C3 RED</b>	<b>C1 GREEN</b>	
<b>Level 1</b>	21.50 – 28.00	21.50 – 28.00	20.00 – 27.00	15.00 – 23.00	
<b>Level 2</b>	37.00 – 45.00	37.50 – 45.00	37.75 – 43.75	42.75 – 51.25	
<i>ChekMate results are reported as % reflectance and should not be compared to values obtained with PTS Panels® test strips.</i> PTS Diagnostics Customer Service: 877-870-5610   Fax: 317-870-5608   Outside of US: +1-317-870-5610 Email: customerservice@ptsdiagnostics.com					
				PS-000882 Rev. 2	

<b>ChekMate Strips Range Card</b>				<b>Batch #:</b>	<b>0860-31017-01</b>
<i>Please enter ranges listed below on the Results Sheet included with your ChekMate kit.</i>				<b>Expiration date:</b>	<b>2019 05 06</b>
	<b>C1 RED</b>	<b>C2 RED</b>	<b>C3 RED</b>	<b>C1 GREEN</b>	
<b>Level 1</b>	21.50 – 28.00	21.50 – 28.00	20.00 – 27.00	15.00 – 23.00	
<b>Level 2</b>	37.00 – 45.00	37.50 – 45.00	37.75 – 43.75	42.75 – 51.25	
<i>ChekMate results are reported as % reflectance and should not be compared to values obtained with PTS Panels® test strips.</i> PTS Diagnostics Customer Service: 877-870-5610   Fax: 317-870-5608   Outside of US: +1-317-870-5610 Email: customerservice@ptsdiagnostics.com					
				PS-000882 Rev. 2	

<b>ChekMate Strips Range Card</b>				<b>Batch #:</b>	<b>0860-31017-01</b>
<i>Please enter ranges listed below on the Results Sheet included with your ChekMate kit.</i>				<b>Expiration date:</b>	<b>2019 05 06</b>
	<b>C1 RED</b>	<b>C2 RED</b>	<b>C3 RED</b>	<b>C1 GREEN</b>	
<b>Level 1</b>	21.50 – 28.00	21.50 – 28.00	20.00 – 27.00	15.00 – 23.00	
<b>Level 2</b>	37.00 – 45.00	37.50 – 45.00	37.75 – 43.75	42.75 – 51.25	
<i>ChekMate results are reported as % reflectance and should not be compared to values obtained with PTS Panels® test strips.</i> PTS Diagnostics Customer Service: 877-870-5610   Fax: 317-870-5608   Outside of US: +1-317-870-5610 Email: customerservice@ptsdiagnostics.com					
				PS-000882 Rev. 2	