

pts panels[®] Multi-Chemistry Controls Range Card		Level 1		Level 2	
Lot #:	MC40	mg/dL	mmol/L	mg/dL	mmol/L
Expiration Date:	2022-03-31				
Single-Analyte Test Strips					
Cholesterol		110-220	2.85-5.70	130-310	3.37-8.03
HDL Cholesterol* <i>(CardioChek[®] Plus analyzer version 1.09 or above)</i>		≤25-60	≤0.65-1.55	30-80	0.78-2.07
Triglycerides		80-200	0.90-2.26	150-300	1.70-3.39
Glucose/eGLU		40-130	2.22-7.22	200-480	11.10-26.64
Multi-Analyte Test Strips					
Cholesterol		105-225	2.72-5.83	120-310	3.11-8.03
HDL Cholesterol* <i>(CardioChek[®] Plus analyzer version 1.09 or above)</i>		≤15-60	≤0.39-1.55	25-85	0.65-2.20
Triglycerides		70-210	0.79-2.37	140-320	1.58-3.62
Glucose		20-135	1.11-7.49	150-420	8.33-23.31
* For CardioChek, CardioChek PA, and CardioChek Plus (version 1.08 or below) analyzers, HDL Cholesterol requires a separate HDL Cholesterol Control (Cat. No. 722).					
PTS Diagnostics Customer Service: 877-870-5610 Fax: 317-870-5608 Outside of US: +1-317-870-5610 Email: customerservice@ptsdiagnostics.com					
PS-002519 Rev. 12					